

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0092
Location:	VERONA POLICE DEPT.		
Calibration File No.:	00702	Calib. Date:	08/25/2021
Certification File No.:	00703	Cert. Date:	08/25/2021
Linearity File No.:	00704	Lin. Date:	08/25/2021
Solution File No.:	00718	Soln. Date:	01/27/2022
Sequential File No.:	00718	File Date:	01/27/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0065
Control Solution %:	0.100%			Expires:	10/21/2022
Solution Control Lot:	20460			Bottle No.:	1291

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:16S		
Control 1 EC	0.100%	03:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:17S		
Control 2 EC	0.099%	03:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18S		
Control 3 EC	0.100%	03:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARATTINI JR

First Name: H.

MI: -

Signature: _____



Badge No.: 9942

Date: 01/27/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 02/08/2022
Certification File No.:	00721	Calib. No.: 00044
Linearity File No.:	00722	Cert. Date: 02/08/2022
Solution File No.:	00726	Cert. No.: 00034
Sequential File No.:	00726	Lin. Date: 02/08/2022
		Lin. No.: 00034
		Soln. Date: 03/03/2022
		Soln. No.: 00216
		File Date: 03/03/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1349

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:52S		
Control 1 EC	0.100%	10:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53S		
Control 2 EC	0.099%	10:54S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54S		
Control 3 EC	0.099%	10:55S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DOWJ P2-210 F.H.

Changed By:

Last Name: HURTADO

First Name: FRANCISCO

MI: A.

Signature: TPH. II Jim Hurtado #7635

Badge No.: 7635

Date: 03/03/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 02/08/2022
Certification File No.:	00721	Calib. No.: 00044
Linearity File No.:	00722	Cert. Date: 02/08/2022
Solution File No.:	00728	Cert. No.: 00034
Sequential File No.:	00728	Lin. Date: 02/08/2022
		Lin. No.: 00034
		Soln. Date: 04/03/2022
		Soln. No.: 00217
		File Date: 04/03/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1205

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:33S		
Control 1 EC	0.099%	01:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:34S		
Control 2 EC	0.098%	01:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:35S		
Control 3 EC	0.099%	01:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 04/03/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 02/08/2022
Certification File No.:	00721	Calib. No.: 00044
Linearity File No.:	00722	Cert. Date: 02/08/2022
Solution File No.:	00731	Cert. No.: 00034
Sequential File No.:	00731	Lin. Date: *02/08/2022
		Lin. No.: 00034
		Soln. Date: 05/02/2022
		Soln. No.: 00218
		File Date: 05/02/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1339

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:22D		
Control 1 EC	0.100%	01:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:23D		
Control 2 EC	0.099%	01:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24D		
Control 3 EC	0.100%	01:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 05/02/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00720	Calib. Date: 02/08/2022
Certification File No.:	00721	Calib. No.: 00044
Linearity File No.:	00722	Cert. Date: 02/08/2022
Solution File No.:	00736	Cert. No.: 00034
Sequential File No.:	00736	Lin. Date: 02/08/2022
		Lin. No.: 00034
		Soln. Date: 06/02/2022
		Soln. No.: 00219
		File Date: 06/02/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1254

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:13D		
Control 1 EC	0.100%	03:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:14D		
Control 2 EC	0.099%	03:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		
Control 3 EC	0.100%	03:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 06/02/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00740	Calib. Date: 07/01/2022
Certification File No.:	00741	Calib. No.: 00045
Linearity File No.:	00742	Cert. Date: 07/01/2022
Solution File No.:	00745	Cert. No.: 00035
Sequential File No.:	00745	Lin. Date: 07/01/2022
		Lin. No.: 00035
		Soln. Date: 07/29/2022
		Soln. No.: 00221
		File Date: 07/29/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1295

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:41D		
Control 1 EC	0.100%	00:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:42D		
Control 2 EC	0.100%	00:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:43D		
Control 3 EC	0.099%	00:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 07/29/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00740	Calib. Date: 07/01/2022
Certification File No.:	00741	Calib. No.: 00045
Linearity File No.:	00742	Cert. Date: 07/01/2022
Solution File No.:	00749	Cert. No.: 00035
Sequential File No.:	00749	Lin. Date: 07/01/2022
		Lin. No.: 00035
		Soln. Date: 08/26/2022
		Soln. No.: 00222
		File Date: 08/26/2022

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1301

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:39D		
Control 1 EC	0.100%	01:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:40D		
Control 2 EC	0.099%	01:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:42D		
Control 3 EC	0.099%	01:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

[Handwritten Signature] #9942

Badge No.: 9942

Date: 08/26/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00740	Calib. Date: 07/01/2022
Certification File No.:	00741	Calib. No.: 00045
Linearity File No.:	00742	Cert. Date: 07/01/2022
Solution File No.:	00755	Cert. No.: 00035
Sequential File No.:	00755	Lin. Date: 07/01/2022
		Lin. No.: 00035
		Soln. Date: 10/01/2022
		Soln. No.: 00223
		File Date: 10/01/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 0278

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:13D		
Control 1 EC	0.100%	02:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14D		
Control 2 EC	0.099%	02:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15D		
Control 3 EC	0.099%	02:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 10/01/2022

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00740 Calib. Date: 07/01/2022 Calib. No.: 00045
Certification File No.: 00741 Cert. Date: 07/01/2022 Cert. No.: 00035
Linearity File No.: 00742 Lin. Date: 07/01/2022 Lin. No.: 00035
Solution File No.: 00758 Soln. Date: 10/30/2022 Soln. No.: 00224
Sequential File No.: 00758 File Date: 10/30/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0163

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:50S		
Control 1 EC	0.100%	01:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:51S		
Control 2 EC	0.099%	01:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:52S		
Control 3 EC	0.099%	01:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: 

Badge No.: 9942

Date: 10/30/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00740	Calib. Date: 07/01/2022
Certification File No.:	00741	Calib. No.: 00045
Linearity File No.:	00742	Cert. Date: 07/01/2022
Solution File No.:	00761	Cert. No.: 00035
Sequential File No.:	00761	Lin. Date: 07/01/2022
		Lin. No.: 00035
		Soln. Date: 11/27/2022
		Soln. No.: 00226
		File Date: 11/27/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 0282

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:42S		
Control 1 EC	0.099%	00:43S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44S		
Control 2 EC	0.098%	00:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45S		
Control 3 EC	0.099%	00:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 11/27/2022