

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00686	Calib. Date: 11/30/2020
Certification File No.:	00687	Calib. No.: 00041
Linearity File No.:	00688	Cert. Date: 11/30/2020
Solution File No.:	00692	Cert. No.: 00031
Sequential File No.:	00692	Lin. Date: 11/30/2020
		Lin. No.: 00031
		Soln. Date: 03/01/2021
		Soln. No.: 00201
		File Date: 03/01/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0068
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	1102

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:33S		
Control 1 EC	0.099%	22:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34S		
Control 2 EC	0.099%	22:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35S		
Control 3 EC	0.099%	22:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 03/01/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00693	Calib. Date: 03/19/2021
Certification File No.:	00694	Cert. No.: 00042
Linearity File No.:	00695	Cert. Date: 03/19/2021
Solution File No.:	00697	Lin. No.: 00032
Sequential File No.:	00697	Lin. Date: 03/19/2021
		Soln. No.: 00203
		Soln. Date: 04/27/2021
		File Date: 04/27/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1141

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:46D		
Control 1 EC	0.099%	23:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:47D		
Control 2 EC	0.099%	23:48D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:49D		
Control 3 EC	0.100%	23:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 04/27/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00693	Calib. Date: 03/19/2021
Certification File No.:	00694	Calib. No.: 00042
Linearity File No.:	00695	Cert. Date: 03/19/2021
Solution File No.:	00698	Cert. No.: 00032
Sequential File No.:	00698	Lin. Date: 03/19/2021
		Lin. No.: 00032
		Soln. Date: 05/30/2021
		Soln. No.: 00204
		File Date: 05/30/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1195

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:57D		
Control 1 EC	0.099%	01:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59D		
Control 2 EC	0.099%	01:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:00D		
Control 3 EC	0.099%	02:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:01D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: 

Badge No.: 9942

Date: 05/30/2021

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00693 Calib. Date: 03/19/2021 Calib. No.: 00042
Certification File No.: 00694 Cert. Date: 03/19/2021 Cert. No.: 00032
Linearity File No.: 00695 Lin. Date: 03/19/2021 Lin. No.: 00032
Solution File No.: 00699 Soln. Date: 07/03/2021 Soln. No.: 00205
Sequential File No.: 00699 File Date: 07/03/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
Control Solution %: 0.100% Expires: 10/21/2022
Solution Control Lot: 20460 Bottle No.: 1298

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:49D		
Control 1 EC	0.099%	21:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50D		
Control 2 EC	0.098%	21:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:52D		
Control 3 EC	0.099%	21:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature:



Badge No.: 9942

Date: 07/03/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00693	Calib. Date: 03/19/2021
Certification File No.:	00694	Calib. No.: 00042
Linearity File No.:	00695	Cert. Date: 03/19/2021
Solution File No.:	00700	Cert. No.: 00032
Sequential File No.:	00700	Lin. Date: 03/19/2021
		Lin. No.: 00032
		Soln. Date: 08/04/2021
		Soln. No.: 00206
		File Date: 08/04/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1299

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:28D		
Control 1 EC	0.099%	04:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:29D		
Control 2 EC	0.099%	04:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:31D		
Control 3 EC	0.099%	04:31D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 08/04/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00702	Calib. Date: 08/25/2021
Certification File No.:	00703	Calib. No.: 00043
Linearity File No.:	00704	Cert. Date: 08/25/2021
Solution File No.:	00710	Cert. No.: 00033
Sequential File No.:	00710	Lin. Date: 08/25/2021
		Lin. No.: 00033
		Soln. Date: 11/01/2021
		Soln. No.: 00210
		File Date: 11/01/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1246

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:51S		
Control 1 EC	0.100%	03:51S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:52S		
Control 2 EC	0.100%	03:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53S		
Control 3 EC	0.100%	03:54S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 11/01/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00702	Calib. Date: 08/25/2021
Certification File No.:	00703	Calib. No.: 00043
Linearity File No.:	00704	Cert. Date: 08/25/2021
Solution File No.:	00713	Cert. No.: 00033
Sequential File No.:	00713	Lin. Date: 08/25/2021
		Lin. No.: 00033
		Soln. Date: 12/26/2021
		Soln. No.: 00212
		File Date: 12/26/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	20460	Expires: 10/21/2022
		Bottle No.: 1347

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:25S		
Control 1 EC	0.099%	23:25S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26S		
Control 2 EC	0.100%	23:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:27S		
Control 3 EC	0.100%	23:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: _____

Badge No.: 9942

Date: 12/26/2021