

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0092  
 Location: VERONA POLICE DEPT.  
 Calibration File No.: 00670 Calib. Date: 01/29/2020 Calib. No.: 00038  
 Certification File No.: 00671 Cert. Date: 01/29/2020 Cert. No.: 00029  
 Linearity File No.: 00672 Lin. Date: 01/29/2020 Lin. No.: 00029  
 Solution File No.: 00675 Soln. Date: 02/03/2020 Soln. No.: 00192  
 Sequential File No.: 00675 File Date: 02/03/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068  
 Control Solution %: 0.100% Expires: 09/09/2021  
 Solution Control Lot: 19240 Bottle No.: 1101

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:34S		
Control 1 EC	0.098%	15:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	15:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:35S		
Control 2 EC	0.096%	15:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37S		
Control 3 EC	0.096%	15:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:38S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 02/03/2020



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0092  
 Location: VERONA POLICE DEPT.  
 Calibration File No.: 00680 Calib. Date: 06/01/2020 Calib. No.: 00040  
 Certification File No.: 00681 Cert. Date: 06/01/2020 Cert. No.: 00030  
 Linearity File No.: 00682 Lin. Date: 06/01/2020 Lin. No.: 00030  
 Solution File No.: 00685 Soln. Date: 11/01/2020 Soln. No.: 00197  
 Sequential File No.: 00685 File Date: 11/01/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065  
 Control Solution %: 0.100% Expires: 09/09/2021  
 Solution Control Lot: 19240 Bottle No.: 1113

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:39S		
Control 1 EC	0.099%	02:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41S		
Control 2 EC	0.098%	02:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:42S		
Control 3 EC	0.098%	02:42S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: \_\_\_\_\_



Badge No.: 9942

Date: 11/01/2020