

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00552	Calib. Date: 10/11/2018
Certification File No.:	00553	Calib. No.: 00033
Linearity File No.:	00554	Cert. Date: 10/11/2018
Solution File No.:	00566	Cert. No.: 00024
Sequential File No.:	00566	Lin. Date: 10/11/2018
		Lin. No.: 00024
		Soln. Date: 01/18/2019
		Soln. No.: 00171
		File Date: 01/18/2019

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXA S3-0065
Control Solution %:	0.100%		Expires: 05/10/2020
Solution Control Lot:	18150		Bottle No.: 0143

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:33S		
Control 1 EC	0.101%	01:34S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:35S		
Control 2 EC	0.099%	01:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:36S		
Control 3 EC	0.099%	01:37S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____



Badge No.: 9942

Date: 01/18/2019

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00552 Calib. Date: 10/11/2018 Calib. No.: 00033
Certification File No.: 00553 Cert. Date: 10/11/2018 Cert. No.: 00024
Linearity File No.: 00554 Lin. Date: 10/11/2018 Lin. No.: 00024
Solution File No.: 00571 Soln. Date: 02/15/2019 Soln. No.: 00172
Sequential File No.: 00571 File Date: 02/15/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0598

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:30S		
Control 1 EC	0.101%	00:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:31S		
Control 2 EC	0.099%	00:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33S		
Control 3 EC	0.099%	00:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:34S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 02/15/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00552	Calib. Date: 10/11/2018
Certification File No.:	00553	Calib. No.: 00033
Linearity File No.:	00554	Cert. Date: 10/11/2018
Solution File No.:	00575	Cert. No.: 00024
Sequential File No.:	00575	Lin. Date: 10/11/2018
		Lin. No.: 00024
		Soln. Date: 03/11/2019
		Soln. No.: 00173
		File Date: 03/11/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0776

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:29S		
Control 1 EC	0.099%	00:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30S		
Control 2 EC	0.100%	00:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32S		
Control 3 EC	0.100%	00:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 03/11/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
 Location: VERONA POLICE DEPT.
 Calibration File No.: 00576 Calib. Date: 03/12/2019 Calib. No.: 00034
 Certification File No.: 00577 Cert. Date: 03/12/2019 Cert. No.: 00025
 Linearity File No.: 00578 Lin. Date: 03/12/2019 Lin. No.: 00025
 Solution File No.: 00588 Soln. Date: 04/07/2019 Soln. No.: 00175
 Sequential File No.: 00588 File Date: 04/07/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
 Control Solution %: 0.100% Expires: 05/10/2020
 Solution Control Lot: 18150 Bottle No.: 0198

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:53S		
Control 1 EC	0.099%	00:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:55S		
Control 2 EC	0.098%	00:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:56S		
Control 3 EC	0.098%	00:57S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:58S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 04/07/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00576	Calib. Date: 03/12/2019
Certification File No.:	00577	Calib. No.: 00034
Linearity File No.:	00578	Cert. Date: 03/12/2019
Solution File No.:	00594	Cert. No.: 00025
Sequential File No.:	00594	Lin. Date: 03/12/2019
		Lin. No.: 00025
		Soln. Date: 05/04/2019
		Soln. No.: 00176
		File Date: 05/04/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0418

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:56D		
Control 1 EC	0.100%	21:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:57D		
Control 2 EC	0.100%	21:57D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58D		
Control 3 EC	0.100%	21:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

[Handwritten Signature]

#9942

Badge No.: 9942

Date: 05/04/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00576 Calib. Date: 03/12/2019 Calib. No.: 00034
Certification File No.: 00577 Cert. Date: 03/12/2019 Cert. No.: 00025
Linearity File No.: 00578 Lin. Date: 03/12/2019 Lin. No.: 00025
Solution File No.: 00601 Soln. Date: 06/01/2019 Soln. No.: 00177
Sequential File No.: 00601 File Date: 06/01/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 1289

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:30D		
Control 1 EC	0.100%	03:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32D		
Control 2 EC	0.100%	03:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		
Control 3 EC	0.100%	03:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____



Badge No.: 9942

Date: 06/01/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00576	Calib. Date: 03/12/2019
Certification File No.:	00577	Calib. No.: 00034
Linearity File No.:	00578	Cert. Date: 03/12/2019
Solution File No.:	00606	Cert. No.: 00025
Sequential File No.:	00606	Lin. Date: 03/12/2019
		Lin. No.: 00025
		Soln. Date: 06/29/2019
		Soln. No.: 00179
		File Date: 06/29/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 1333

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:58D		
Control 1 EC	0.099%	22:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:59D		
Control 2 EC	0.098%	23:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:01D		
Control 3 EC	0.099%	23:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

[Handwritten Signature] #9942

Badge No.: 9942

Date: 06/29/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00576	Calib. Date: 03/12/2019
Certification File No.:	00577	Calib. No.: 00034
Linearity File No.:	00578	Cert. Date: 03/12/2019
Solution File No.:	00605	Cert. No.: 00025
Sequential File No.:	00605	Lin. Date: 03/12/2019
		Lin. No.: 00025
		Soln. Date: 06/29/2019
		Soln. No.: 00178
		File Date: 06/29/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 1292

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:09D		
Control 1 EC	----%	21:10D	25.3°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	21:10D	25.3°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature:

Badge No.: 9942

Date: 06/29/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00576	Calib. Date: 03/12/2019
Certification File No.:	00577	Calib. No.: 00034
Linearity File No.:	00578	Cert. Date: 03/12/2019
Solution File No.:	00610	Cert. No.: 00025
Sequential File No.:	00610	Lin. Date: 03/12/2019
		Lin. No.: 00025
		Soln. Date: 07/27/2019
		Soln. No.: 00180
		File Date: 07/27/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 1317

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:19D		
Control 1 EC	0.096%	22:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20D		
Control 2 EC	0.097%	22:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:22D		
Control 3 EC	0.097%	22:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 07/27/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0092
Location:	VERONA POLICE DEPT.		
Calibration File No.:	00619	Calib. Date:	08/28/2019
Certification File No.:	00620	Cert. No.:	00037
Linearity File No.:	00621	Lin. Date:	08/28/2019
Solution File No.:	00626	Cert. No.:	00028
Sequential File No.:	00626	Lin. No.:	00028
		Soln. Date:	09/23/2019
		Soln. No.:	00183
		File Date:	09/23/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0065
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	1339

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:42D		
Control 1 EC	0.099%	00:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:44D		
Control 2 EC	0.099%	00:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:45D		
Control 3 EC	0.100%	00:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

HERIBERTO CARATTINI JR.

Badge No.: 9942

Date: 09/23/2019

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00619 Calib. Date: 08/28/2019 Calib. No.: 00037
Certification File No.: 00620 Cert. Date: 08/28/2019 Cert. No.: 00028
Linearity File No.: 00621 Lin. Date: 08/28/2019 Lin. No.: 00028
Solution File No.: 00645 Soln. Date: 11/02/2019 Soln. No.: 00185
Sequential File No.: 00645 File Date: 11/02/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 1327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:34S		
Control 1 EC	0.100%	23:35S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:35S		
Control 2 EC	0.099%	23:36S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37S		
Control 3 EC	0.098%	23:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:38S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 11/02/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00619	Calib. Date: 08/28/2019
Certification File No.:	00620	Cert. No.: 00037
Linearity File No.:	00621	Cert. Date: 08/28/2019
Solution File No.:	00655	Lin. No.: 00028
Sequential File No.:	00655	Lin. Date: 08/28/2019
		Soln. No.: 00187
		Soln. Date: 11/15/2019
		File Date: 11/15/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 1293

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:50S		
Control 1 EC	0.100%	22:51S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:52S		
Control 2 EC	0.099%	22:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53S		
Control 3 EC	0.099%	22:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 11/15/2019

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00619	Calib. Date: 08/28/2019
Certification File No.:	00620	Calib. No.: 00037
Linearity File No.:	00621	Cert. Date: 08/28/2019
Solution File No.:	00663	Cert. No.: 00028
Sequential File No.:	00663	Lin. Date: 08/28/2019
		Lin. No.: 00028
		Soln. Date: 12/08/2019
		Soln. No.: 00188
		File Date: 12/08/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1104

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:47S		
Control 1 EC	0.099%	00:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:48S		
Control 2 EC	0.097%	00:49S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:50S		
Control 3 EC	0.097%	00:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.	First Name: HERIBERTO	MI: -
Signature: 	Badge No.: 9942	Date: 12/08/2019

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00619	Calib. Date: 08/28/2019
Certification File No.:	00620	Calib. No.: 00037
Linearity File No.:	00621	Cert. Date: 08/28/2019
Solution File No.:	00666	Cert. No.: 00028
Sequential File No.:	00666	Lin. Date: 08/28/2019
		Lin. No.: 00028
		Soln. Date: 12/14/2019
		Soln. No.: 00189
		File Date: 12/14/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	19240	Expires: 09/09/2021
		Bottle No.: 1218

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:27S		
Control 1 EC	0.099%	00:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:29S		
Control 2 EC	0.097%	00:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30S		
Control 3 EC	0.097%	00:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 12/14/2019