

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00490 Calib. Date: 10/24/2017 Calib. No.: 00031
Certification File No.: 00491 Cert. Date: 10/24/2017 Cert. No.: 00022
Linearity File No.: 00492 Lin. Date: 10/24/2017 Lin. No.: 00022
Solution File No.: 00515 Soln. Date: 01/18/2018 Soln. No.: 00154
Sequential File No.: 00515 File Date: 01/18/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 0586

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:20S		
Control 1 EC	0.101%	04:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:22S		
Control 2 EC	0.099%	04:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:23S		
Control 3 EC	0.099%	04:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:24S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 01/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00490	Calib. Date: 10/24/2017
Certification File No.:	00491	Cert. No.: 00031
Linearity File No.:	00492	Cert. Date: 10/24/2017
Solution File No.:	00522	Lin. No.: 00022
Sequential File No.:	00522	Lin. Date: 10/24/2017
		Soln. No.: 00157
		Soln. Date: 02/18/2018
		File Date: 02/18/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0143

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:56S		
Control 1 EC	0.099%	14:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 2 EC	0.098%	14:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		
Control 3 EC	0.099%	15:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUFF

First Name: EDWARD

MI: H

Signature: _____



Badge No.: 2912

Date: 02/18/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00490 Calib. Date: 10/24/2017 Calib. No.: 00031
Certification File No.: 00491 Cert. Date: 10/24/2017 Cert. No.: 00022
Linearity File No.: 00492 Lin. Date: 10/24/2017 Lin. No.: 00022
Solution File No.: 00524 Soln. Date: 03/17/2018 Soln. No.: 00158
Sequential File No.: 00524 File Date: 03/17/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 0236

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:39S		
Control 1 EC	0.101%	01:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:41S		
Control 2 EC	0.099%	01:41S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:42S		
Control 3 EC	0.099%	01:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:43S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 03/17/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00531	Calib. Date: 04/13/2018
Certification File No.:	00532	Calib. No.: 00032
Linearity File No.:	00533	Cert. Date: 04/13/2018
Solution File No.:	00535	Cert. No.: 00023
Sequential File No.:	00535	Lin. Date: 04/13/2018
		Lin. No.: 00023
		Soln. Date: 05/18/2018
		Soln. No.: 00160
		File Date: 05/18/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:59D		
Control 1 EC	0.098%	04:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:01D		
Control 2 EC	0.097%	04:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	04:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:02D		
Control 3 EC	0.096%	04:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature:

Badge No.: 9942

Date: 05/18/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00531	Calib. Date: 04/13/2018
Certification File No.:	00532	Calib. No.: 00032
Linearity File No.:	00533	Cert. Date: 04/13/2018
Solution File No.:	00536	Cert. No.: 00023
Sequential File No.:	00536	Lin. Date: 04/13/2018
		Lin. No.: 00023
		Soln. Date: 06/15/2018
		Soln. No.: 00161
		File Date: 06/15/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0013

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:27D		
Control 1 EC	0.099%	04:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:29D		
Control 2 EC	0.099%	04:29D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:30D		
Control 3 EC	0.098%	04:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 06/15/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00531 Calib. Date: 04/13/2018 Calib. No.: 00032
Certification File No.: 00532 Cert. Date: 04/13/2018 Cert. No.: 00023
Linearity File No.: 00533 Lin. Date: 04/13/2018 Lin. No.: 00023
Solution File No.: 00538 Soln. Date: 07/12/2018 Soln. No.: 00162
Sequential File No.: 00538 File Date: 07/12/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
Control Solution %: 0.100% Expires: 10/17/2019
Solution Control Lot: 17360 Bottle No.: 1343

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:51D		
Control 1 EC	0.097%	03:51D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:52D		
Control 2 EC	0.096%	03:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53D		
Control 3 EC	0.096%	03:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:54D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINIR,

First Name: HERIBERTO

MI: -

Signature:  #9942

Badge No.: 9942

Date: 07/12/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00531	Calib. Date: 04/13/2018
Certification File No.:	00532	Calib. No.: 00032
Linearity File No.:	00533	Cert. Date: 04/13/2018
Solution File No.:	00543	Cert. No.: 00023
Sequential File No.:	00543	Lin. Date: 04/13/2018
		Lin. No.: 00023
		Soln. Date: 08/17/2018
		Soln. No.: 00164
		File Date: 08/17/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	17360	Expires: 10/17/2019
		Bottle No.: 0550

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:06D		
Control 1 EC	0.095%	04:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	04:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:07D		
Control 2 EC	0.095%	04:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	04:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08D		
Control 3 EC	0.095%	04:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	04:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

Badge No.: 9942

Date: 08/17/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00531 Calib. Date: 04/13/2018 Calib. No.: 00032
Certification File No.: 00532 Cert. Date: 04/13/2018 Cert. No.: 00023
Linearity File No.: 00533 Lin. Date: 04/13/2018 Lin. No.: 00023
Solution File No.: 00546 Soln. Date: 09/02/2018 Soln. No.: 00165
Sequential File No.: 00546 File Date: 09/02/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0308

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:49D		
Control 1 EC	0.095%	02:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	02:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:50D		
Control 2 EC	0.095%	02:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51D		
Control 3 EC	0.096%	02:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	02:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: _____

 #9942

Badge No.: 9942

Date: 09/02/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00531	Calib. Date: 04/13/2018
Certification File No.:	00532	Calib. No.: 00032
Linearity File No.:	00533	Cert. Date: 04/13/2018
Solution File No.:	00550	Cert. No.: 00023
Sequential File No.:	00550	Lin. Date: 04/13/2018
		Lin. No.: 00023
		Soln. Date: 09/29/2018
		Soln. No.: 00166
		File Date: 09/29/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0068
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0645

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:58D		
Control 1 EC	0.096%	10:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.095%	11:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	11:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		
Control 3 EC	0.095%	11:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature: 

Badge No.: 9942

Date: 09/29/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00552	Calib. Date: 10/11/2018
Certification File No.:	00553	Calib. No.: 00033
Linearity File No.:	00554	Cert. Date: 10/11/2018
Solution File No.:	00556	Cert. No.: 00024
Sequential File No.:	00556	Lin. Date: 10/11/2018
		Lin. No.: 00024
		Soln. Date: 10/26/2018
		Soln. No.: 00168
		File Date: 10/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0837

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:44D		
Control 1 EC	0.101%	02:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 2 EC	0.100%	02:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46D		
Control 3 EC	0.100%	02:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature:

Badge No.: 9942

Date: 10/26/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00552	Calib. Date: 10/11/2018
Certification File No.:	00553	Calib. No.: 00033
Linearity File No.:	00554	Cert. Date: 10/11/2018
Solution File No.:	00560	Cert. No.: 00024
Sequential File No.:	00560	Lin. Date: 10/11/2018
		Lin. No.: 00024
		Soln. Date: 11/23/2018
		Soln. No.: 00169
		File Date: 11/23/2018

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0065
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 0778

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:50S		
Control 1 EC	0.100%	08:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		
Control 2 EC	0.099%	08:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52S		
Control 3 EC	0.099%	08:53S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUFF

First Name: EDWARD

MI: H

Signature:  #2912

Badge No.: 2912

Date: 11/23/2018

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0092
Location:	VERONA POLICE DEPT.	
Calibration File No.:	00552	Calib. Date: 10/11/2018
Certification File No.:	00553	Cert. No.: 00033
Linearity File No.:	00554	Cert. Date: 10/11/2018
Solution File No.:	00563	Lin. No.: 00024
Sequential File No.:	00563	Soln. Date: 12/20/2018
		Soln. No.: 00170
		File Date: 12/20/2018

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXA S3-0065
Control Solution %:	0.100%		Expires: 05/10/2020
Solution Control Lot:	18150		Bottle No.: 0360

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:56S		
Control 1 EC	0.101%	03:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57S		
Control 2 EC	0.099%	03:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:59S		
Control 3 EC	0.100%	03:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: HERIBERTO

MI: -

Signature:

Badge No.: 9942

Date: 12/20/2018