

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00810 Calib. Date: 09/29/2023 Calib. No.: 00048
Certification File No.: 00811 Cert. Date: 09/29/2023 Cert. No.: 00038
Linearity File No.: 00812 Lin. Date: 09/29/2023 Lin. No.: 00038
Solution File No.: 00827 Soln. Date: 01/26/2024 Soln. No.: 00243
Sequential File No.: 00827 File Date: 01/26/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0068
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0214

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:45S		
Control 1 EC	0.102%	01:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:47S		
Control 2 EC	0.100%	01:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:48S		
Control 3 EC	0.100%	01:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:50S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature: 

Badge No.: 9942

Date: 01/26/2024

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00830 Calib. Date: 02/06/2024 Calib. No.: 00049
Certification File No.: 00831 Cert. Date: 02/06/2024 Cert. No.: 00039
Linearity File No.: 00832 Lin. Date: 02/06/2024 Lin. No.: 00039
Solution File No.: 00837 Soln. Date: 03/03/2024 Soln. No.: 00245
Sequential File No.: 00837 File Date: 03/03/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 0399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:14S		
Control 1 EC	0.099%	00:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:15S		
Control 2 EC	0.098%	00:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:17S		
Control 3 EC	0.098%	00:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CARATTINI JR.

First Name: H.

MI: -

Signature:



Badge No.: 9942

Date: 03/03/2024

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0092
Location: VERONA POLICE DEPT.
Calibration File No.: 00830 Calib. Date: 02/06/2024 Calib. No.: 00049
Certification File No.: 00831 Cert. Date: 02/06/2024 Cert. No.: 00039
Linearity File No.: 00832 Lin. Date: 02/06/2024 Lin. No.: 00039
Solution File No.: 00841 Soln. Date: 04/01/2024 Soln. No.: 00246
Sequential File No.: 00841 File Date: 04/01/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0065
Control Solution %: 0.100% Expires: 01/09/2026
Solution Control Lot: 24020 Bottle No.: 1039

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:45S		
Control 1 EC	0.100%	00:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47S		
Control 2 EC	0.099%	00:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:48S		
Control 3 EC	0.099%	00:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CARATTINI JR

First Name: H.

MI: -

Signature: _____



Badge No.: 9942

Date: 04/01/2024